



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 1:55 pm, Apr 03, 2013

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201195	NAME OF AGENCY Grandview PD	DATE OF INSPECTION 04/02/2013
LOCATION OF INSTRUMENT (STREET AND CITY) 1200 Main St, Grandview MO. 64056		TIME OF INSPECTION 8:13 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 04/02/13 08:13
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER
<input checked="" type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER Guth Labs LOT # 11220 EXP. DATE 10/24/2013	
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN DR5385 EXP. DATE 12/10/2013	
<input checked="" type="checkbox"/> CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT) Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED) <input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE <input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE <input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE	

TEST 1 .102%	TEST 2 .102%	TEST 3 .102%
--------------	--------------	--------------

☒ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 10	(0-.04) 4	(.05-.09) 1	(.10-.14) 2	(.15-.19) 2	OVER .19 1
-------------	-----------	-------------	-------------	-------------	------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS
(USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME Ryan A Sharp
TYPE II PERMIT NUMBER/EXPIRATION DATE 220074 03/08/2014	TELEPHONE NUMBER (816) 316-4900

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **11220** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 27, 2011**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1205%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 24, 2013** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN102408-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



RYAN SHARP

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 03/08/2012

Number 220074

Expires 03/08/2014

MO 580-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)

BAC DataMaster

Evidence Ticket

4444. (E) 11/15/2014
 4445. (E) 11/15/2014

2008 年 10 月 10 日 星期五 09:00
 10/10/2008
 10:00

... *Adipobutyrin* 100 (1933) 38

Abstract:

11000000 000 000 00000000 00000000

NAME: _____
 STUDENT NUMBER: _____

11:46 AM - 12:00 PM

CONCLUSION: The results of this study suggest that the use of a single, low-dose, short-acting benzodiazepine may be a safe and effective method of sedation for the placement of a nasogastric tube in the emergency department.

66-67-12	5	0000000000	000000
----------	---	------------	--------

1996. 2000. 2004. 2008. 2012. 2016. 2020.	1996. 2000. 2004. 2008. 2012. 2016. 2020.
---	---

MINI REVIEW

¹ "The National Commission on the Causes and Prevention of Violence," *Report of the National Commission on the Causes and Prevention of Violence* (New York: Random House, 1969), pp. 1-2.

ATOR SIGNATURE

Stock No.

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI
COUNTY OF [COUNTY NAME]

UNITED STATES DISTRICT COURT
SOUTHERN DISTRICT OF NEW YORK

[illegible]

647X(2) 1.011(2) Å,

2000年10月10日星期二

63. NAME DATE GRADE

INSTRUMENTAL APPROACHES TO THE

[illegible][illegible]

OPERATOR SIGNATURE

Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI
GRANDVIEW POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201195
04/02/13

ARREST TIME: 07:00

SUBJECT NAME:

SMITH

DOB: 11/11/11

SEX: M

STATE/D.L.: MO-12345

ARRESTING OFFICER:

SHARP

OFFICER I.D.: 147

TESTING OFFICER:

SHARP

OFFICER I.D.: 147

PERMIT NUMBER: 220674

EXPIRATION DATE: 03/08/14

MISCELLANEOUS DATA:

BREATH ANALYSIS

BLANK TEST

.000

08:06

INTERNAL STANDARD

VERIFIED

08:06

RADIO INTERFERENCE

OPERATOR SIGNATURE

Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901